

Porcine Serum

Certificate of Analysis

0.2 µm sterile filtered
Manufactured in EU

Lot No: 32112320PS
Product Code: S-PS-EU-015, S-PS-EU-011
Origin: EU
Date of manufacture: 11/2020
Expiry Date: 11/2025
Storage: < -20°C

Physical and Chemical Analysis

| Test | Specifications | Results |
|------------|-------------------|-------------|
| pH | 7.0 – 8.0 | 7.55 |
| Endotoxin | < 50 EU/mL | 26.28 EU/mL |
| Osmolality | 260 – 350 mOsm/kg | 291 mOsm/kg |
| Hemoglobin | < 50 mg/dL | 38 mg/dL |

Protein Profile

| Test | Specifications | Results |
|---------------|----------------|-----------|
| Total Protein | 6.0 – 8.0 g/dL | 6.85 g/dL |

Sterility Testing

| Test | Specifications | Results |
|-----------|----------------|---------|
| Sterility | Pass | Pass |

India Contact:
Life Technologies (India) Pvt. Ltd.
Ph: +91-11-42208000, 42208111, 42208222, Mobile: +91-9810521400
Email: customerservice@lifetechindia.com Website: www.lifetechindia.com

CAUTION: THIS PRODUCT IS NOT INTENDED FOR HUMAN OR ANIMAL CONSUMPTION OR THERAPEUTIC USE.